

POX VACCINATOR PARTS LIST

- 31 Syringe body
- 32 End front cap
- 33 Rubber stopper
- 34 Neoprene washer
- 35 Needle
- 36 Needle screw
- 38 Glass barrel
- 311- Needle shaft
- 313- Body cover
- 315- Handle spring
- 317- Upper handle
- 333- Spare parts kit (33, 34, 38)






Automatic Veterinary Syringes

Automatic Fowl Pox Vaccinator

The Automatic Fowl Pox Vaccinator is made of the best materials, is oil-proof, non corrosive and all parts can be sterilized



Operating Instructions:

- 1. Open the end front cap (No. 32) of the Vaccinator.
- 2. Fill the vaccine direct to the glass barrel.
- 3. Close the end front cap (No. 32) tight to the glass barrel.
- 4. Every squeezing on the upper handle, bring a drop direct to the chicken wing.
- 5. After use; open the end front cap dismantle the glass barrel and flush the complete vaccinator with clear water.
- 6. Sterilize; you may sterilize the vaccinator parts up to 125°C before the next use.